

**REGULATORY GUIDE B7**

**COMPLYING WITH TITLE B - MAMMOGRAPHY**



SC DEPARTMENT *of*  
**ENVIRONMENTAL  
SERVICES**

**September 2025**

# Table of Contents

FACILITY REGISTRATION APPROVAL.....	3
SHIELDING REQUIREMENTS.....	3
REGISTERING EQUIPMENT .....	4
FACILITY CERTIFICATION .....	4
EQUIPMENT .....	4
TECHNOLOGISTS .....	5
INTERPRETING PHYSICIANS .....	5
MEDICAL PHYSICIST .....	6
QUALITY ASSURANCE REQUIREMENTS (EQUIP - ENHANCING QUALITY USING THE INSPECTION PROGRAM).....	7
MEDICAL RECORDS AND MAMMOGRAPHY REPORTS.....	8
MAMMOGRAPHY MEDICAL OUTCOMES AUDIT .....	10
EQUIPMENT EVALUATIONS .....	10
ANNUAL SURVEY .....	11
QUALITY CONTROL .....	11
SELF-REFERRALS .....	12
INSPECTIONS .....	12
QUESTIONS .....	12
REGULATORY GUIDES.....	13

## **REGULATORY GUIDE B7**

### **COMPLYING WITH TITLE B – MAMMOGRAPHY**

Each mammography facility that is registered with the Department is required to comply with Regulations 61-64, X-Rays (Title B), which are the regulations concerning x-ray equipment. This guide is intended to assist the mammography facility in complying with Title B regulations.

#### **FACILITY REGISTRATION APPROVAL**

(See RHB2.4)

Prior to installing an x-ray machine, a facility must apply to the Department for a Facility Registration Approval (FRA). To receive a Facility Registration Approval, complete and return SCDES Form 0845 "Facility Registration Approval Request" along with the non-refundable application fee of \$62.50. This form may be submitted by email to XRayFRA-ShieldingPlans@des.sc.gov or mailed to 2600 Bull Street, Columbia, SC 29201. Fees may be paid by check, or an invoice will be issued after the application is received. Applications are processed in the order received, with a typical turnaround time of 15 business days (excluding weekends and holidays) once a complete submission and payment are received. As the Department transitions to electronic correspondence, all approval letters and documentation will be delivered electronically. Please maintain digital records, as paper copies will not be mailed.

#### **A FACILITY SHALL NOT INSTALL AN X-RAY PRODUCING MACHINE UNTIL THE DEPARTMENT HAS ISSUED A FACILITY REGISTRATION APPROVAL.**

If a facility relocates, SCDES Form 4262 "Report of Change" must be submitted to the Department within thirty (30) calendar days. This form may be submitted by email to XRayRegistration@des.sc.gov or mailed to 2600 Bull Street, Columbia, SC 29201. Facility Registration Approval is not transferable to a new owner or additional locations. A new Facility Registration Approval and associated fees are required for the acquisition of an existing facility.

Please note that certain types of equipment require a shielding plan prior to installation in accordance with RHB 4.4. See Regulatory Guide B6 or contact the Department for assistance.

#### **SHIELDING REQUIREMENTS**

(See RHB 4.4)

A shielding plan or radiation area survey, which is acceptable only if prior approval is given, is required for mammography units. Shielding plans or a written request to perform a post-install survey in lieu of a shielding plan must be submitted to and accepted by this Department prior to installation. Shielding plans and area surveys must be prepared and submitted by a registered vendor (Class III or IX).

To request review, complete and submit SCDES Form 0846 "Shielding Plan Review Application," the shielding plan, and supporting documents along with the non-refundable \$62.50 fee. Both shielding plans and requests for post installation radiation area surveys require the submission of a shielding review fee of \$62.50. Submit by email to XRayFRA-ShieldingPlans@des.sc.gov or by mail to 2600 Bull Street, Columbia, SC 29201. Fees may be paid by check, or an invoice will be issued after the application is received. Applications are processed in the order received, with a typical turnaround time of 15 business days (excluding weekends and holidays) once a

complete submission and payment are received.

### **REGISTERING EQUIPMENT**

(See RHB 2.5)

All x-ray equipment must be registered with the Department within thirty (30) calendar days of installation by submitting SCDES Form 0819 "Registration and/or Deletion of X-ray Producing Machines." This form may be submitted by email to XRayRegistration@des.sc.gov or mailed to 2600 Bull Street, Columbia, SC 29201. The Department will issue a registration sticker for each control panel. Stickers must be affixed in a clearly visible location to indicate the facility's registration status.

See Regulatory Guide B1 for additional details on registration requirements.

### **FACILITY CERTIFICATION**

(See RHB 5.2)

Certification is a process that is administered by SCDES/FDA. SCDES issues an MQSA certificate upon notification by the ACR that a mammography facility has been accredited.

Certification is valid for three years and can be renewed as long as the facility remains properly accredited and successfully demonstrates during its annual inspections that it continues to meet the MQSA quality standards.

Only SCDES/FDA-certified facilities can lawfully provide mammography services. The Centers for Medicare and Medicaid Services (CMS) will reimburse only for the mammography performed at a SCDES/FDA-certified facility.

### **EQUIPMENT**

(See RHB 5.8)

The facility should have documentation showing that each unit has been accredited by the accreditation body or, for new units, showing that the unit has passed an equipment evaluation or medical physicist's survey and that the application for accreditation of the unit has been submitted.

There are four cases where the units in use in the facility may not need to be accredited:

- 1) The unit is a "loaner" while repairs to the facility's unit are taking place (limited to 30 days without extenuating documentation), \*
- 2) The unit is installed in the facility for evaluation prior to purchase (limited to not more than 90 days), \*
- 3) If the unit is used for interventional mammography only or, \*
- 4) The unit is an experimental one installed and used under an Investigational Device Exemption (IDE) as described in the Safe Medical Devices Act of 1990 or other FDA approved research protocol.

\*Note that under 1), 2) and 3) the unit still must have passed an equipment evaluation or survey, and each such unit will be tested by the state inspector, during the annual inspection, regardless of its accreditation or ownership status.

In all cases the shielding must be accepted prior to use and the unit must be registered with this department within thirty (30) days after installation of the unit.

## **TECHNOLOGISTS**

(See RHB 5.9.2)

1. In order to independently perform mammographic examinations, one must meet the following requirements:
  - a) Be certified and maintain a valid certificate issued by the South Carolina Radiation Quality Standards Association to perform general radiographic procedures. For more information, contact the SCRQSA at (877)771-6141 or visit their website at [www.scrqsa.org](http://www.scrqsa.org).
  - b) Have prior to April 28, 1999, qualified as a radiologic technologist under the interim MQSA regulations; OR completed 40 contact hours of specific training in mammography in the topics specified in the regulations, including performance of a minimum of 25 examinations under direct supervision.
2. After meeting all the initial requirements, all radiologic technologists must:
  - a) Maintain mammography continuing education (15 CEU's/36 months).
  - b) Maintain continuing experience (200 examinations/24 months).
  - c) Requalification: Radiologic technologists failing to maintain the continuing requirements must requalify prior to independently performing mammographic examinations.
    - Continuing Education: Must bring total up to 15 CEU's/36 months.
    - Continuing Experience: Perform a minimum of 25 mammography examinations under direct supervision. At the end of the 6-month grace period, the technologist must have performed 200 examinations.
  - d) Maintain a valid certificate to perform general radiographic procedures from the South Carolina Radiation Quality Standards
3. Must have 8 Hours of training in each modality used by the technologist, i.e. Full Field Digital Mammography (FFDM), Digital Breast Tomosynthesis (DBT) systems, etc.

## **INTERPRETING PHYSICIANS**

(See RHB 5.9.1)

1. In order to independently interpret mammograms, one must qualify as an interpreting physician. To do this, one must have either qualified as an interpreting physician under the interim regulations prior to April 28, 1999, OR after April 28, 1999, have documented all of the following requirements:
  - a) Have a valid South Carolina State license to practice medicine.
  - b) Be Board Certified in Diagnostic Radiology by an FDA-approved body or have 3 months of formal training in mammography.
  - c) Have 60 Category I CME credits in mammography with at least 15 obtained in the 3 years immediately prior to qualifying as an interpreting physician.

- d) Have interpreted, under direct supervision, the mammographic examinations from 240 patients in the 6 months immediately prior to qualifying as an interpreting physician OR if the physician passed their certifying board in diagnostic radiology at the first allowable time the 6 month period could have been anytime in the last two years of the residency program.
2. After meeting all the initial requirements, all interpreting physicians must:
- a) Maintain mammography continuing education (15 Category I CME's/36 months).
  - b) Maintain continuing experience (960 examinations/24months).
  - c) Physicians failing to maintain the continuing requirements must requalify prior to performing independent mammographic interpretation.
    - Continuing education: Must bring total up to 15 CME's/3 years.
    - Continuing experience: Interpret 240 examinations under direct supervision or interpret a sufficient number, under direct supervision, to bring total to 960/24 months, whichever is less. At the end of the 6-month grace period, the physician must have interpreted at least 960 examinations.
  - d) Maintain a valid South Carolina State license to practice medicine.
3. Must have 8 Hours of training in each modality used by the interpreting physician, i.e. Full Field Digital Mammography (FFDM), Digital Breast Tomosynthesis (DBT) systems, etc.

### **MEDICAL PHYSICIST**

(See RHB 5.9.3)

In order to independently conduct surveys of mammography facilities and provide oversight of a facility's quality assurance program, one must qualify as a medical physicist. Because the duties of the medical physicist encompass more than just the physics survey, FDA expects the facility to be able to call on the services of the medical physicist throughout the year. Therefore, the facility must be able to identify their qualified medical physicist at the time of the DES inspection.

1. To qualify under the initial qualifications of the final regulations, a medical physicist must document all of the following requirements:
- a) Be registered as a Class IX vendor.
  - b) Have a master's degree or higher in a physical science with no less than 20 semester hours in physics.
  - c) Have 20 contact hours of specialized training in conducting mammography facility surveys.
  - d) Have the experience of conducting surveys of at least 1 mammography facility and a total of at least 10 mammography units.

OR

To qualify under the alternative initial qualifications of the final regulations, a medical physicist must document all of the following requirements:

- a) Be registered as a Class IX vendor.
  - b) Have qualified as a medical physicist under the interim regulations and maintained active status of any licensure, approval, or certification required under the interim regulations.
  - c) Prior to April 28, 1999, have:
    - A bachelor's degree or higher in a physical science with no less than 10 semester hours in physics.
    - Forty contact hours of documented specialized training in conducting surveys of mammography facilities and,
    - Have the experience of conducting surveys of at least 1 mammography facility and a total of at least 20 mammography units. The training and experience requirements must be met after fulfilling the degree requirement.
2. Having once met the initial qualifications, all medical physicists (whether they initially qualified under the interim or final regulations) must meet the continuing qualifications:
- a) Continuing education (15 CEU's/36 months).
  - b) Continuing experience (2 facilities and 6 units/24 months).
  - c) Maintain Class IX registration.
  - d) Medical physicists failing to maintain the continuing requirements must requalify prior to independently conducting surveys of mammography facilities.
    - Continuing education: Must bring total up to 15 CEU's/3 years.
    - Continuing experience: Must bring total up to 2 facilities and 6 units/24 months under direct supervision.
  - e) Must have 8 Hours of training in each modality surveyed by the physicist, i.e. Full Field Digital Mammography (FFDM), Digital Breast Tomosynthesis (DBT) systems, etc.

#### **QUALITY ASSURANCE REQUIREMENTS (EQUIP - ENHANCING QUALITY USING THE INSPECTION PROGRAM)**

(See RHB 5.12)

Since the quality of mammograms is one of the most important determinants of the accuracy of mammography, the production of high quality clinical images by certified mammography facilities is one of the primary goals of the Mammography Quality Standards Act (MQSA).

All interpreting physicians shall participate in quality assurance activities as described in 5.12.1.2.

Each mammographic facility shall designate one interpreting physician to be the lead interpreting physician (LIP), who is responsible for ensuring that the facility's quality assurance program meets the requirements of these regulations.

All interpreting physicians (IP) for the facility are required to follow the facility procedures for corrective action when the images that they are asked to interpret are of poor quality. The facility must have a mechanism for the IP to provide feedback to the RT's or other designated facility personnel when images

are of poor quality. The facility must have a mechanism to document corrective action taken and the effectiveness of the corrective action.

Facilities must have a system in place to ensure that images continue to comply with the clinical image quality standards established by the facility's accreditation body. The facility must perform regular reviews of image quality attributes of a sample of mammograms performed by each active RT and of mammograms accepted for interpretation by each active IP. During each annual inspection, inspectors will ask for documentation that the facility performed a clinical image review at least once since the last inspection.

The LIP is responsible for providing oversight of the QA and QC records, including a review of the frequency of performance of all required tests, and review of any corrective actions when needed. The LIP must be either available to answer questions on the day of the inspection or sign an attestation provided to the facility.

In addition, all interpreting physicians for the facility are required to participate in the facility's medical outcomes audit program. Each mammographic facility shall also designate at least one interpreting physician to be the reviewing interpreting physician(s), who is/are responsible for analyzing the medical outcomes audit. The reviewing interpreting physician may or may not be the lead interpreting physician.

Physicians in training may work at facilities as long as they are under the direct supervision of a qualified interpreting physician.

### **MEDICAL RECORDS AND MAMMOGRAPHY REPORTS**

(See RHB 5.11)

Each facility shall prepare a written report of the results of each mammography examination performed under its certificate. The mammography report shall include the following information:

- 1) The name of the patient and an additional patient identifier
- 2) Date of examination
- 3) The name of the interpreting physician who interpreted the mammogram
- 4) Overall final assessment of findings

The overall final assessment of findings must be classified in one of the following categories:

- 1) **Negative**
- 2) **Benign**
- 3) **Probably Benign**
- 4) **Suspicious**
- 5) **Highly suggestive of malignancy**
- 6) **Incomplete: Need additional imaging evaluation**

While the final assessment findings must not vary from these categories and must be stated as written above, limited flexibility is allowed for further description as long as it doesn't change the meaning of the category. The following are considered equivalent to the wording listed above and are acceptable final overall assessments:



- 1) **Negative**  
*Negative Mammogram*
- 2) **Benign**  
*Benign Finding*  
*Benign Findings*  
*Benign Abnormality*  
*Benign Abnormalities*  
*Benign Mammogram*
- 3) **Probably Benign**  
*Probably Benign Finding*  
*Probably Benign Findings*  
*Probably Benign Abnormality*  
*Probably Benign Abnormalities*  
*Probably Benign-Short Interval Follow-up Suggested*  
*Probably Benign Finding-Short Interval Follow-up Suggested*  
*Probably Benign Mammogram*
- 4) **Suspicious**  
*Suspicious Finding*  
*Suspicious Findings*  
*Suspicious Abnormality*  
*Suspicious Abnormalities*  
*Suspicious for Malignancy*  
*Suspicious Abnormality-Biopsy Should Be Considered*  
*Suspicious Finding-Biopsy Should Be Considered*  
*Suspicious Mammogram*
- 5) **Highly Suggestive of Malignancy**  
*Highly Suggestive of Malignancy*  
*Highly Suggestive of Malignance - Appropriate Action should be Taken*
- 6) **Incomplete: Need Additional Imaging Evaluation**  
*Incomplete: Needs Additional Imaging Evaluation*  
*Incomplete: Additional Imaging Evaluation Needed*  
*Incomplete: Need Additional Imaging Evaluation -Comparison with Prior Studies*  
*Incomplete: Need additional imaging evaluation and/or prior mammograms for comparison*  
*Incomplete: Need prior mammograms for comparison*  
*Need Additional Imaging Evaluation (the term "Incomplete" can be inferred in this example as this is the only incomplete BIRADS assessment category)*  
*Incomplete Mammogram: Need Additional Imaging Evaluation*

Each facility shall send each patient a summary of the mammography report written in lay terms within 30 days of the mammographic examination. If assessments are "Suspicious" or "Highly suggestive of malignancy", the facility shall make reasonable attempts to ensure that the results are communicated to the patient as soon as possible. Each facility that accepts patients who do not have a health care provider shall maintain a system for referring such patients to a health care provider when clinically indicated.

When the patient has a referring health care provider or the patient has named a health care provider, the

facility shall provide a written report of the mammography examination to that health care provider as soon as possible, but no later than 30 days after the date of the examination. If the assessment is "Suspicious" or "Highly suggestive of malignancy", the facility must make reasonable attempts to communicate with the health care provider as soon as possible, or if the health care provider is unavailable, to a responsible designee of the health care provider.

Each facility that performs mammograms shall maintain mammography films and reports in a permanent record of the patient for a period of not less than 5 years, or not less than 10 years if no additional mammograms of the patient are performed at the facility. Upon request by, or on behalf of, the patient, the facility shall permanently or temporarily transfer the original mammograms and copies of the patient's reports to a medical institution, or to a physician or health care provider of the patient or to the patient directly. Any fee charged to the patient for providing films and/or reports shall not exceed the documented costs associated with this service.

### **MAMMOGRAPHY MEDICAL OUTCOMES AUDIT**

(See RHB 5.23)

Each facility shall establish and maintain a mammography medical outcomes audit program to follow-up positive mammography assessments and to correlate pathology results with the interpreting physician's findings. This program shall be designed to ensure the reliability, clarity and accuracy of the interpretation of mammograms.

Each facility shall establish a system to collect and review outcome data for all mammograms performed, including follow-up on the disposition of all positive mammograms and correlation of pathology results with the interpreting physician's mammography report. Analysis of these outcome data shall be made individually and collectively for all interpreting physicians at the facility. In addition, any cases of breast cancer among patients imaged at the facility that subsequently become known to the facility shall prompt the facility to initiate follow-up on surgical and/or pathology results and review of the mammograms taken prior to the diagnosis of a malignancy.

The facility's first audit analysis shall be initiated no later than twelve months after the date the facility becomes certified. This audit analysis shall be completed within an additional twelve months to permit completion of diagnostic procedures and data collection. Subsequent audit analyses shall be conducted at least once every twelve months.

Each facility shall designate at least one interpreting physician to review the medical outcomes audit data at least once every twelve months. This individual shall record the dates of the audit period(s) and shall be responsible for analyzing results based on this audit. The individual shall also be responsible for documenting the results, notifying other interpreting physicians of their results and the facility results, as a whole. If follow-up actions are taken, the reviewing interpreting physician shall also be responsible for documenting the nature of the follow-up.

### **EQUIPMENT EVALUATIONS**

(See RHB 5.15)

Whenever a new unit or processor is installed, disassembled and reassembled at the same or a new location, or major components are changed or repaired, an evaluation of the mammography unit or image processor is required. You must refer to your manufacturer's quality control manual to determine whether or not the

change requires a mammography equipment evaluation.

This additional evaluation is needed to verify that all functions that may have been affected by the change or repair have been successfully restored even if a full survey had recently been completed. For a new unit, an equipment evaluation is needed before the unit is used on patients unless the unit has already undergone a full survey. In this situation, the facility must follow the accreditation body procedures. Keep in mind that under MQSA, the facility has the ultimate responsibility for ensuring image quality and patient safety.

The equipment evaluation must be performed by a qualified medical physicist or by an individual under the direct supervision of the medical physicist. These evaluations will be used to determine whether the new or changed equipment meets the requirements of applicable standards. All problems shall be corrected before the new or changed equipment is put into service for examinations or image interpretation. A facility should maintain the mammography equipment evaluation report and documentation of any corrective action until the next inspection that verifies compliance.

### **ANNUAL SURVEY**

(See RHB 5.14)

1. At least once a year, each facility shall undergo a survey by a medical physicist or by an individual under the direct supervision of a medical physicist. At a minimum, this survey shall include the performance of tests to ensure that the facility meets the quality assurance requirements of the annual tests described in parts 5.13.5 and 5.13.6 or 5.13.7; and the weekly phantom image quality test described in part 5.13.2.
2. The results of all tests conducted by the facility in accordance with part 5.13.1, as well as written documentation of any corrective actions taken, and their results shall be evaluated for adequacy by the medical physicist performing the survey.
3. The medical physicist shall prepare a survey report that includes a summary of this review and recommendations for necessary improvements.
4. The survey report shall be sent to the facility within 30 days of the date of the survey.
5. The survey report shall be dated and signed by the medical physicist performing or supervising the survey. If the survey was performed entirely or in part by another individual under the direct supervision of the medical physicist, that individual and the part of the survey that individual performed shall also be identified in the survey report.

### **QUALITY CONTROL**

(See RHB 5.13.7)

Facilities should follow the quality control program recommended by the manufacturer of their mammography unit. Average glandular dose must meet the requirements of part 5.13.5.10.

It is the mammography facility's responsibility to ensure that the required quality control testing is being performed for all off-site review workstations. Facilities should regularly obtain quality control records for offsite RWS's and maintain the records for inspection by the Department.

The manufacturer's current operating manual shall be available for Department review.

## **SELF-REFERRALS**

(See Appendix A)

Self-referrals cannot be accepted without prior approval of Appendix A (Information to be Submitted by Persons Proposing to Conduct Healing Arts Screening) from Title B. Once approved, any changes, such as equipment replacement, a change in film or screens, or new personnel, must be reported to the Department within fifteen (15) days.

## **INSPECTIONS**

(See RHB 1.7)

The Department conducts annual inspections of mammography facilities. The Department will also conduct inspections if a valid complaint is received. If violations are reported on an inspection, a follow-up inspection may be conducted if the severity of the violations warrants it. Generally, the Department will call a facility a minimum of five days in advance to schedule an inspection. Inspections by the Department are mandatory, but every attempt will be made to accommodate patient schedules. **The Department does have the right to make unannounced inspections.**

The inspection consists of checking the operation of the x-ray equipment, as well as checking administrative items such as records. Generally, an inspection requires use of the mammography equipment for about fifteen minutes. The facility can greatly assist the Department inspector by reviewing and completing the Inspection Confirmation information sheet which is sent to the facility once the inspection has been scheduled.

At the conclusion of the inspection, the inspector will conduct an exit interview to discuss any items of noncompliance, as well as any other pertinent observations. The inspector will mail the facility an inspection report documenting any violations or comments.

If the MQSA (Mammography Quality Standards Act) report results in a Level 1, repeat Level 1 or repeat Level 2 finding, a written Corrective Action Plan shall be provided to the Department within 15 calendar days of the date of citation. If the MQSA report results in a Level 2 or repeat Level 3 finding, a written Corrective Action Plan shall be provided to the Department within 30 calendar days of the date of citation.

The facility must notify the Department, in writing, by the appropriate date that corrections have been made. Corrective action must be described for each violation. Each violation must be addressed individually. It will not suffice to simply state that all violations have been corrected. After the Department has received and accepted the required corrective action, a Completed Corrective Action letter will be sent to the facility.

## **QUESTIONS**

If you have questions, please feel free to call or write:

SC Department of Environmental Services  
Radiation Protection Program  
2600 Bull Street

Columbia, SC 29201  
PHONE: (803) 545-4400  
FAX: (803) 545-4412

EMAIL –

REGISTRATIONS: XRayRegistration@des.sc.gov

FACILITY REGISTRATION APPROVALS AND SHIELDING PLANS: XRayFRA-ShieldingPlans@des.sc.gov

### **REGULATORY GUIDES**

- B1 - Registration of X-ray Facilities and Equipment
- B2 - Complying with Title B - Medical Facilities
- B3 - Complying with Title B - Dental Facilities
- B4 - Complying with Title B - Facilities Utilizing Analytical or Industrial X-ray Equipment
- B5 - Vendor Registration and Responsibilities
- B6 - X-Ray Facility Shielding Plans
- B7 - Complying with Title B - Mammography
- B8 - Complying with Title B - Personnel Security Screening Systems
- B9 - Complying with Title B - Veterinary Facilities
- B10 - Complying with Title B - Hospitals

**Visit our website at: <https://des.sc.gov/programs/bureau-land-waste-management/radiation-protection-program/x-ray-regulations>**